



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

**Bureau Veritas**  
2450 Hassell Road  
Hoffman Estates, IL 60195

Fulfills the requirements of

**ISO/IEC 17025:2017**

In the field of

**TESTING**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

A handwritten signature in black ink, appearing to be 'Jason Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 23 December 2027

Certificate Number: L2264.03



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**Bureau Veritas**  
 2450 Hassell Road  
 Hoffman Estates, IL 60195  
 Jennifer Anweiler  
 Quality & Technical Manager  
 (800) 655-4473

**TESTING**

ISO/IEC 17025 Accreditation Granted: **22 December 2025**

Certificate Number: **L2264.03**      Certificate Expiry Date: **23 December 2027**

**Chemical**

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Accelerated Stability	OCM SOP-00029	Fuel	Oil bath/Filtration
Total Acid Number	ASTM D664 Mod	Oil	Manual Titration
Total Acid Number	ASTM D974 Mod	Oil, Fuel	Manual Titration
API Gravity	ASTM D287	Oil, Fuel	Hydrometer
Appearance	ASTM D4176	Fuel	Visual
Physical Properties of Coolants	OCM SOP-00007	Coolant	Visual
Ash Content	ASTM D482	Fuel	Hot plate/open flame
Microorganisms	OCM SOP-00022	Any Fluid	Slide
Total Base Number	ASTM D4739 Mod	Oil	Manual Titration
Total Base Number	OCM-SOP-00017 TBN	Oil	Manual Titration
Biodiesel Content by FTIR	LOC WI-00018	Fuel	FTIR
Carbon Residue	ASTM D524	Fuel	Furnace
Cetane Index, Calculated	ASTM D976 ASTM D4737	Fuel	Calculation
Chlorides	LOC WI-00032	MWF	Test Strip

This Scope of Accreditation, version 009, was last updated on: 22 December 2025 and is valid only when accompanied by the Certificate.

Page 1 of 4

**Chemical**

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Cloud Point	ASTM D2500	Fuel	Cloud Point Unit
Cold Flow Plug Point	ASTM D6371	Fuel	CFPP Unit
Color, ASTM	ASTM D1500	Any Fluid	Color Comparator
Color, Gardner and Sayboldt	ASTM D1500	Oil	Color Comparator
Conductivity	LOC WI-00030	MWF	Conductivity Meter
Copper Strip Corrosion	ASTM D130	Oil, Fuel	Copper Specimen
Density Specific Gravity	ASTM D1298	Oil, Fuel	Hydrometer
Distillation	ASTM D86	Fuel	Distillation
Ferroggraphy DR	OCM SOP-00021	Oil	DR3 Unit
ELC Coolant Inhibitor	OCM SOP-00008	Coolant	Test Strip
Flash Point Closed Cup	ASTM D93	Oil, Fuel	Automated Flash Point
Flash, Point Open Cup	ASTM D92	Oil	Automated Flash Point
Freeze Point of Coolant	ASTM D3321 (Mod)	Coolant	Refractometer
FTIR Infrared Analysis	OCM – SOP-0005-FTIR	Any Fluid	FTIR
Fuel Dilution by GC	ASTM D7593	Oil	GC
Glycol	OCM SOP-00004	Oil	Wet Chemistry
Estimate of Net & Gross Heat of Combustion Fuel	ASTM D4868	Fuel	Calculation
Heat of Combustion of Liquid Hydrocarbon Fuels by Bomb Calorimeter	ASTM D4809	Fuel	Bomb Calorimeter
Karl Fischer Water	ASTM D6304	Oil	Karl Fischer Titrator
Lubricity by HFRR	ASTM D6079	Fuel	Compass PCS HFRR Unit
Metals by ICP	ASTM D5185 Mod	Any Fluid	ICP
MWF-Brix	LOC WI-00032	Aqueous fluids & Metal working fluids	Refractometer
MWF- Iron Chip Corrosion	LOC WI-00029	Metal working fluids	Wet Chemistry

This Scope of Accreditation, version 009, was last updated on: 22 December 2025 and is valid only when accompanied by the Certificate.

Page 2 of 4

**Chemical**

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
MWF- Dirt by Filtration	LOC WI-00031	Metal working fluids	Filtration
MWF- Free Oil	LOC WI-00033	Metal working fluids	Wet Chemistry
MWF- pH & Alkalinity	LOC WI-00034	Metal working fluids	pH Meter
MWF-Tramp Oil	LOC WI-00035	Metal working fluids	Centrifuge
Nitrites	OCM SOP-00023	Coolant	Manual Titration
Particle Count	ASTM D7647	Oil	APC Unit
Particulate Contamination	ASTM D5452 ASTM D6217	Fuel	Filtration
pH of Coolants	ASTM D1287	Coolant	pH Meter
Pour Point	ASTM D97	Fuel	Pour Point Unit
Refrigerants, Testing of Low & High Pressure	LOC WI-00017	Refrigerant	KF Titration
Reserve Alkalinity	ASTM D1121	Coolant	Manual Titration
Sulfated Ash	ASTM D874	Oil, Fuel	Hot Plate/ Open Flame
Sulfur in Petroleum by XRF Spectrometer	ASTM D4294	Oil, Fuel	XRF Spectrometer
Sulfur, Total by UV Fluorescence	ASTM D5453	Fuel	UV Fluorescence
Total Hardness (CaCO <sub>3</sub> )	LOC WI-00032	MWF	ICP
Varnish by Quantitative Spectrophotometric Analysis (QSA)	OCM SOP-00011	Oil	Spectrophotometer
Viscosity	ASTM D445	Oil, Fuel	Viscometer
Viscosity	ASTM D7279 Mod	Oil, Fuel	Houillon Viscometer
Viscosity Index	ASTM D2270	Oil	Calculation
Viscosity, SUS Calculated	ASTM D2161	Any Fluid	Calculation
Water and Sediment	ASTM D2709 ASTM D1796 Mod	Oil, Fuel	Centrifuge
Water Percent Estimate by Hotplate	OCM SOP-00024	Oil	Hot Plate

**Chemical**

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Additives Elements and Wear Metals Contaminants	ASTM D6130 Mod	Coolant	ICP



Jason Stine, Vice President

